

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/UP/26/01)

Certificate of Verification**BHASKAR SCALE ENGINEERING, B398 PATEL NAGAR 2ND GHAZIABAD-201001, Ph: 9871254243****Certificate No:** IND/GATC/UP/26/01/26/75**Date of Verification:** 04-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-M/s **Neeraj Sanitry and Hardware Store**,Address-**Chaproula, Main GT Road, Greater Noida, Gautambuddha Nagar, Uttar Pradesh,Gautam Buddha Nagar,Uttar Pradesh,201009, Ph:- 9810522829**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	BHASKAR	14243	2026	Class III	100kg	200g	10g	kg	10g	10000	15g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	25	50%	240V 50Hz	Yes	INDGATCU P2601B	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	Reverification

Model Approval No(s) - IND/09/19/430**Verification Fee (Rs. + GST) - 1534; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 1534** deposited vide Money Receipt No. **102**, Dated **04-06-2026**Repaired by / Used by: **BHASKAR SCALE ENGINEERING**Next verification falls due on or before: **2027-06-03****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.